

Certificate of Analysis

Product Description: HTLV-II Antibody Positive Plasma Confirmed
 Material No(s): 0325-0032, 0325-0033, 0325-0034, 0325-0035
 Lot No: 9212068
 Donor ID: BD110127
 Gender: NA
 Age: NA
 Country of Origin: NA
 Collection Date: NA
 Anticoagulant: NA
 Storage Temperature: -20 degrees C
 Shipping Conditions: Dry ice
 Intended Use: For research use only. Not for use in diagnostic procedures.
 Safety Precautions: Use universal precautions when handling this material.
 Test Method(s): See below

Lot Analysis					
Test Date	Analyte	Test Method/Manufacturer	Result	Units	Notes
15-Jun-07	HTLV-I/II Antibody	Biomerieux EIA	2.575	S/CO	
21-Sep-07	HTLV-I/II Antibody	Abbott EIA	0.488	S/CO	1:10 dilution
21-Sep-07	HTLV-I/II Antibody	Abbott EIA	0.519	S/CO	1:10 dilution
21-Sep-07	HTLV-I/II Antibody	Abbott EIA	0.315	S/CO	1:20 dilution
21-Sep-07	HTLV-I/II Antibody	Abbott EIA	0.342	S/CO	1:20 dilution
21-Sep-07	HTLV-I/II Antibody	Abbott EIA	2.412	S/CO	
21-Sep-07	HTLV-I/II Antibody	Abbott EIA	2.496	S/CO	
27-Dec-12	HTLV-I/II Antibody	MP Diagnostics EIA	0.602	S/CO	1:10 dilution
27-Dec-12	HTLV-I/II Antibody	MP Diagnostics EIA	6.679	S/CO	
02-Aug-07	HTLV-I/II Western Blot	Genelabs Diagnostics Western Blot	HTLV-II	NA	
02-Aug-07	19 Band	Genelabs Diagnostics Western Blot	-	NA	
02-Aug-07	21 Band	Genelabs Diagnostics Western Blot	+	NA	
02-Aug-07	24 Band	Genelabs Diagnostics Western Blot	+	NA	
02-Aug-07	26 Band	Genelabs Diagnostics Western Blot	-	NA	
02-Aug-07	28 Band	Genelabs Diagnostics Western Blot	-	NA	
02-Aug-07	32 Band	Genelabs Diagnostics Western Blot	-	NA	
02-Aug-07	36 Band	Genelabs Diagnostics Western Blot	-	NA	
02-Aug-07	46 Band	Genelabs Diagnostics Western Blot	-	NA	
02-Aug-07	53 Band	Genelabs Diagnostics Western Blot	-	NA	
02-Aug-07	RI Band	Genelabs Diagnostics Western Blot	-	NA	
02-Aug-07	RII Band	Genelabs Diagnostics Western Blot	+	NA	
This product was tested for HBsAg, HIV Ag, antibody to HCV, and antibody to HIV-1/2 and found to be negative.					

Data are provided for informational purposes. SeraCare Life Sciences does not claim that others can duplicate test results exactly.

NA = Not Available



Certificate of Analysis

Product Description: HTLV-II Antibody Positive Plasma Confirmed
Material No(s): 0325-0032, 0325-0033, 0325-0034, 0325-0035
Lot No: 9212068

Approval:

Brenda de Freitas

15Sep15

Prepared By

Date

Christine Paiva

22Sep15

QA Verified By

Date

