

Certificate of Analysis

Revised 14OCT2020

Product Description: SARS-CoV-2 Antibody Positive Plasma
 Material Number: 0325-0087, 0325-0088, 0325-0089, 0325-0090
 Batch: 10496584
 Donor ID: BD436209
 Gender: Male
 Age: 34
 Country of Origin: United States
 Collection Date: 14MAY2020
 Anticoagulant: 4% Sodium Citrate
 Storage: -20 Degree C
 Shipping Conditions: Dry ice
 Intended Use: For research use or for further manufacture into in-vitro diagnostic reagents.
 Safety Precautions: Use universal precautions when handling this material.
 Test Method(s): See below

Data are provided for informational purposes.

SeraCare Life Sciences does not claim that others can duplicate the test results indicated below.

Batch Analysis					
Test Date	Analyte	Test Method/Manufacturer	Result	Units	Notes
21-May-20	SARS-CoV-2 IgG	Abbott Architect	4.230	S/CO	≥ 1.4=Pos
21-May-20	SARS-CoV-2 IgG	Abbott Architect	1.810	S/CO	1:4 Dilution
23-May-20	SARS-CoV-2 Total	Roche Elecsys	7.190	COI	1:4 Dilution
23-May-20	SARS-CoV-2 Total	Roche Elecsys	46.230	COI	≥ 1.0=Pos
02-Oct-20	SARS-CoV-2 IgM	Beckman Dxl 600	1.570	S/CO	≥ 1.0=Pos
This product was tested for HBsAg, and antibodies to HCV and HIV1/2, HBV/HCV/HIV NAT and Syphilis and found to be negative.					
Additional Data					
Date of Symptom Onset		30Mar2020			
Date of PCR Testing		02Apr2020			
Date of Last Symptom(s)		04Apr2020			

NA = Not Available

Approval:



14 OCT 2020

Prepared By

Date



15 OCT 2020

QA Verified By

Date